

Certificate of Analysis

BiomatiQ™




Effective Date: 23 JUL 2025

Product Name : 10 X 75 mm Endotoxin Free Assay Tubes, Borosilicate Glass USP Type I
Batch No. : D1010072550A **Product Code** : DAP01075-50
Mfg. Month/Year : 07/2025 **Expiry Month/Year** : 06/2027
Report No. : DPN/FG/G25/618 **Released on** : 23/07/2025

S. No.	Test parameter	Specification	Result
1	Appearance	No visible particulate or defects, clean, intact	Complies
2	Endotoxin Level	<0.005 EU/mL	<0.005 EU/mL

- This certifies that the above product has been processed under a validated dry heat Depyrogenation cycle ($\geq 250^{\circ}\text{C}$ for ≥ 90 minutes) and packed under controlled environmental conditions.
- The product meets all quality requirements and is labelled as Endotoxin Free.
- Depyrogenation records are electronically generated

Conclusion: The Batch No. **D1010072550A** **Complies** with the Spec. No. **SPC/FGPQ/DPN007**
Endotoxin Free & tested by Kinetic Chromogenic Assay

Sign & Date	Prepared By (QC)	Reviewed By (QA)	Approved By (Head QA/ Designee)
	 23/07/25	 23/07/25	 23/07/25