

Certificate of Analysis






Effective Date: 31 JUL 2025

Product Name : 16 X 100 mm Endotoxin Free Dilution Tubes, Borosilicate Glass USP Type I
Batch No. : D1615072516A **Product Code** : DAP16100-16
Mfg. Date : 07/2025 **Expiry Date.** : 06/2027
Report No. : DPN/FG/G25/657 **Released on** : 31/07/2025

S. No.	Test parameter	Specification	Result
1	Appearance	No visible particulate or defects, clean, intact	Complies
2	Endotoxin Level	<0.005 EU/mL	<0.005 EU/mL

- This certifies that the above product has been processed under a validated dry heat Depyrogenation cycle ($\geq 250^{\circ}\text{C}$ for ≥ 90 minutes) and packed under controlled environmental conditions.
- The product meets all quality requirements and is labelled as Endotoxin Free.
- Depyrogenation records are electronically generated.

Conclusion: The Batch No. **D1615072516A** Complies with the Spec. No. **SPC/FGPQ/DPN009**. Endotoxin Free & tested by Kinetic Chromogenic Assay.

Sign & Date	Prepared By (QC)	Reviewed By (QA)	Approved By (Head QA/ Designee)
	 31/07/25	 31/07/2025	 31/07/25